Ozcanium Analytics
ABN: 14 754 514 569
Ozcaniumanalytics.com.au
+61 407 260 481

Certificate of Analysis

Kisspeptin 5mg

Client Name : Australian Peptide Labs
Method : Bacterial Endotoxin (LAL)

Batch Number : APL-KISS5-2609-01
Analysis Date : 09/09/2026
Verification Code : OZ-ENDOTX-TEP6

METHOD & INSTRUMENT

Test	Bacterial Endotoxin Test
Assay principle	Kinetic Chromogenic LAL
Instrument	Endosafe nexgen-PTS
Device serial	21461229
Method code	KX-122
Cartridge	Endotoxin-LAL
Cartridge range	5 - 0.05 EU/mL
Test temperature	37.0 °C (start/end 37.0)
Sample dilution	1:100

SAMPLE IMAGE

STANDARD CURVE

Curve slope	-0.352
Curve intercept	+2.419
Reaction time window	149 - 753 s

REPORTED ENDOTOXIN CONCENTRATION
< 15.0 EU/vial

RESULT VALID
System suitability met
Reported at 1:100 dilution

Reporting Notes

-The '<' symbol denotes that endotoxin in the sample is below the reportable value shown, at the dilution tested.
-The reportable value equals the cartridge lower sensitivity (0.05 EU/mL) multiplied by the sample dilution factor (100).
-Endotoxin content is reported for the sample as received.

Analysis by Ibrahim A. Ozcan,
Lead Chemist Ozcanium Analytics
Signature

For research use only, not for human, clinical, therapeutic or manufacturing-release use.
Results relate only to the item tested and this certificate shall not be reproduced

Report Compiled by Zein S. Sabini
09/09/2026

© Copyright 2026 by Ozcanium Analytics. All rights reserved.